



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

LIFELOC TECHNOLOGIES
12441 W. 49th Ave #4
Wheat Ridge, CO 80033
Mark Lary Phone: 303 431 9500

CALIBRATION

Valid To: December 31, 2025

Certificate Number: 5342.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,3}:

I. Chemical

Parameter/Equipment	Range	CMC ² (±)	Comments
Gas Detection Equipment – Ethanol Concentration	(0.008 to 0.099) g/210 L (0.100 to 0.413) g/210 L	0.003 g/210 L 3 % of reading	SI traceable wet bath, ethanol standards g/210 L is breath or international equivalent

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

LIFELOC TECHNOLOGIES

Wheat Ridge, CO

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 18th day of December 2023.

A blue ink signature of Mr. Trace McInturff, written in a cursive style.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 5342.01
Valid to December 31, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.